

Deutsche Akkreditierungsstelle GmbH

Annex to the Accreditation Certificate D-K-15089-01-01 according to DIN EN ISO/IEC 17025:2005

Valid from: 25.07.2019

Date of issue: 25.07.2019

Holder of certificate:

Perschmann Calibration GmbH
Hauptstraße 46d, 38110 Braunschweig

Head:

Deputy head:

Dr. Detlef Rübesame
Dipl.-Wirtsch.-Ing. Lars Ahrendt
Heiner Klimke
Dipl.-Ing. (FH) Stefan Bosse
Klaus Wedekind
Andreas Brech
Bernd Wierzbowski
Thomas Zelm
Thomas Grobe
Marco Hirsch

Accredited as calibration laboratory since: 30.07.1993

Annex to the accreditation certificate D-K-15089-01-01

Calibration in the fields:

Dimensional quantities

Length

- Gauge blocks
- Length measuring instruments
- Line scales, distances
- Length measuring devices ^{a)}
- Diameter
- Form error
- Flatness ^{a)}
- Straightness ^{a)}
- Thread

Coordinate measuring technology

- Coordinate measuring machines ^{c)}

Mechanical quantities

- Torque ^{b)}
- Pressure
- Weighing instruments ^{a)}
- Material testing machines (MTM)
 - Hardness (MTM)

Electrical quantities

DC and low frequency

- DC voltage
- AC voltage
- DC current
- AC current
- DC resistance

Time and frequency

- Frequency

Thermodynamic quantities

Temperature quantities

- Temperature indicators and simulators

^{a)} also on-site calibration

^{b)} also on-site calibration and calibration in the mobile laboratory

^{c)} only on-site calibration

Within the measurands/calibration items marked with with *, the calibration laboratory is permitted, without being required to inform and obtain prior approval from DAkkS, to use calibration standards or equivalent calibration procedures listed here with different issue dates.

The calibration laboratory maintains a current list of all calibration standards / equivalent calibration procedures within the flexible scope of accreditation.

Abbreviations used: see last page

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Length				
Gauge blocks * made of steel according to DIN EN ISO 3650:1999	0.5 mm to 100 mm featuring the nominal values of the standards	DAkkS-DKD-R 4-3 part 3.1:2010 Measurement of the deviation of the central length l_m from the nominal value l_n by comparison measurement Measurement of the deviations f_0 and f_u from the central length by 5 points comparison For the smallest meas- urement uncertainties, the wringability and the wringing characteristics of both measuring surfaces must be checked using an appropriate optical flat	For the central length: $0.05 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot l$ For the deviations f_0 and f_u from the central length: $0.05 \mu\text{m}$	Measuring surface quality as stated in QMH resp. in the test specifications l = gauge block length
	> 100 mm to 150 mm featuring the nominal values of the standards		For the central length: $0.05 \mu\text{m} + 0.7 \cdot 10^{-6} \cdot l$ For the deviations f_0 and f_u from the central length: $0.07 \mu\text{m}$	
Gauge blocks * made of ceramics according to DIN EN ISO 3650:1999	0.5 mm to 100 mm featuring the nominal values of the standards		For the central length: $0.07 \mu\text{m} + 0.6 \cdot 10^{-6} \cdot l$ For the deviations f_0 and f_u from the central length: $0.05 \mu\text{m}$	
	> 100 mm to 150 mm featuring the nominal values of the standards		For the central length: $0.07 \mu\text{m} + 0.8 \cdot 10^{-6} \cdot l$ For the deviations f_0 and f_u from the central length: $0.07 \mu\text{m}$	
Gauge blocks * made of steel with special cross section (round or square), also with drilling in the middle	0.5 mm to 100 mm		For the central length: $0.1 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot l$ For the deviations f_0 and f_u from the central length: $0.1 \mu\text{m}$	At square gauge blocks with drillings the mean size is substituted by ANSI-ASME B89.1.9M measured between hole and front side
Gauge blocks * made of tungsten carbide according to DIN EN ISO 3650:1999	0.5 mm to 100 mm featuring the nominal values of the standards		For the central length: $0.08 \mu\text{m} + 1.2 \cdot 10^{-6} \cdot l$ For the deviations f_0 and f_u from the central length: $0.05 \mu\text{m}$	
Gauge blocks * made of steel	> 150 mm to 1000 mm in the nominal dimensions, which differ of the stan- dard with a max. of 50 mm	Measurement of the deviation of the central length l_c from the nominal value l_n by comparison measurement	For the central length: $0.2 \mu\text{m} + 0.7 \cdot 10^{-6} \cdot l$	

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-01

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Setting ring gauges * made of steel Diameter	2 mm to 200 mm	DAkkS-DKD-R 4-3 part 4.1:2010	$0.4 \mu\text{m} + 2 \cdot 10^{-6} \cdot d$	The measurement uncertainty applies to the complete calibration of diameter, roundness, straightness and parallelism. For the calibration of the diameter without form measurement, the best measurement uncertainty increases by $0.2 \mu\text{m}$ d = measured diameter
Setting plug gauges * made of steel Diameter	1 mm to 200 mm	DAkkS-DKD-R 4-3 part 4.1:2010	$0.4 \mu\text{m} + 2 \cdot 10^{-6} \cdot d$	
Measuring pins * made of steel Diameter	0.17 mm to 20 mm	DAkkS-DKD-R 4-3 part 4.2:2010	$0.4 \mu\text{m}$	
Roundness deviation * of abovementioned rings, inside cylinders, plugs or outside cylinders	to 40 μm	DAkkS-DKD-R 4-3 part 4.1:2010	$0.2 \mu\text{m} + 1 \cdot 10^{-2} \cdot R_{\text{ONT}}$	Diameter: 2 mm to 200 mm
Straightness deviation * of abovementioned rings, inside cylinders, plugs or outside cylinders	to 10 μm	DAkkS-DKD-R 4-3 part 4.1:2010	$0.5 \mu\text{m}$	axial length: to 30 mm
Setting dimension *	25 mm to 900 mm	VDI/VDE/DGQ 2618 part 4.4:2009	$0.7 \mu\text{m} + 1,5 \cdot 10^{-6} \cdot l$	l = measured length
Caliper gauge *	5 mm to 170 mm	DAkkS-DKD-R 4-3 part 4.7:2010	$1.5 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Reference- and Setting gauge	to 12 mm	Annex F/43:2017-11	$3 \mu\text{m}$	Feeler gauge, gap gauge of plastic, delrin, teflon, brass or steel
Radius gauge	to 40 mm	Annex F/42:2018-01	$3 \mu\text{m}$	to 40 mm radii
Angle meter *	0° to 360°	DAkkS-DKD-R 4-3 part 7.2:2010	$1' 30''$	
Graduator	0° to 180°	Annex F/46:2017-11	$12'$	
Measuring tape Circumference tape measure	0 m to 50 m	Annex F/47-1:2017-12 Annex F/47-2:2017-12	$50 \mu\text{m} + 15 \cdot 10^{-6} \cdot l$	l = measured length
Diameter tape measure	0 m to 10 m	Annex F/47-2:2017-12	$50 \mu\text{m} + 15 \cdot 10^{-6} \cdot d$	d = measured diameter
Ruler	0 m to 10 m	Annex F/47-3:2017-12 Annex F/47-4:2017-11	$50 \mu\text{m} + 15 \cdot 10^{-6} \cdot l$	l = measured length Graduated metal rules, reference- and plotting scale, rules, folding rules
Calipers for external, internal and depth dimensions *	0 mm to 1000 mm	DAkkS-DKD-R 4-3 part 9.1:2010	$30 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	l = measured length
Depth calipers *	0 mm to 1000 mm	DAkkS-DKD-R 4-3 part 9.2:2010	$30 \mu\text{m} + 30 \cdot 10^{-6} \cdot l$	

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-01

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range		Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Height gauge *	0 mm	to 1000 mm	DAkks-DKD-R 4-3 part 9.3:2010	$20\text{ }\mu\text{m} + 10 \cdot 10^{-6} \cdot l$	with contact help
Micrometers *	0 mm	to 100 mm	DAkks-DKD-R 4-3 part 10.1:2010	$3\text{ }\mu\text{m} + 10 \cdot 10^{-6} \cdot l$	100 mm = final value of the measuring range
Indicating caliper micrometers *	0 mm	to 100 mm	DAkks-DKD-R 4-3 part 10.3:2010	$2\text{ }\mu\text{m} + 10 \cdot 10^{-6} \cdot l$	l = measured length
Indicating caliper gap gauge	0 mm	to 100 mm	Annex F/39:2017-12	$2\text{ }\mu\text{m} + 10 \cdot 10^{-6} \cdot l$	
Depth micrometers *	0 mm	to 300 mm	VDI/VDE/DGQ 2618 part 10.5:2010	$3\text{ }\mu\text{m} + 10 \cdot 10^{-6} \cdot l$	with interchangeable contact points
Internal micrometers with two-point contact *	25 mm	to 950 mm	DAkks-DKD-R 4-3 part 10.7:2010	$3.5\text{ }\mu\text{m} + 5 \cdot 10^{-6} \cdot d$	d = measured diameter
Internal micrometers with jaws	5 mm	to 100 mm	Annex F/37:2017-11	$5\text{ }\mu\text{m} + 5 \cdot 10^{-6} \cdot d$	
Internal micrometers with three-point contact *	3 mm	to 200 mm	DAkks-DKD-R 4-3 part 10.8:2010	$3\text{ }\mu\text{m} + 5 \cdot 10^{-6} \cdot d$	
Dial gauges with scales * Scale interval > 1 μm		to 100 mm	DAkks-DKD-R 4-3 part 11.1:2010	$3\text{ }\mu\text{m} + 10 \cdot 10^{-6} \cdot l$	l = measured length
Dial gauges with scales * Scale interval 1 μm		to 5 mm		$1.5\text{ }\mu\text{m}$	error of measurement y_i
				$2\text{ }\mu\text{m}$	deviation span f_e, f_{ges}, f_u, f_t and f_w
Dial gauges with scales * Scale interval > 1 μm		to 100 mm	VDI/VDE/DGQ 2618 part 11.1:2014	$3\text{ }\mu\text{m} + 10 \cdot 10^{-6} \cdot l$	l = measured length
Dial gauges with scales * Scale interval 1 μm		to 5 mm		$1.5\text{ }\mu\text{m}$	error of measurement y_i
				$2\text{ }\mu\text{m}$	deviation span $MPE_r, MPE_e, MPE_{ges}, MPE_{1/1}, MPE_{1/2}, MPE_{1/10}, MPE_u$
Dial gauges with digital display Numerical interval 0.1 μm		to 25 mm	Annex F/04-2 V5:2014	$0.6\text{ }\mu\text{m} + 10 \cdot 10^{-6} \cdot l$	error of measurement y_i
				$0.8\text{ }\mu\text{m} + 10 \cdot 10^{-6} \cdot l$	deviation span f_e, f_t and f_w
Dial gauges with digital display Numerical interval 1 μm		to 100 mm		$1\text{ }\mu\text{m} + 10 \cdot 10^{-6} \cdot l$	error of measurement y_i
				$1.5\text{ }\mu\text{m} + 10 \cdot 10^{-6} \cdot l$	deviation span f_e, f_t and f_w
Dial indicators * Scale interval > 0.5 μm		to 3 mm	DAkks-DKD-R 4-3 part 11.2:2010	0.6 μm	
Lever gauges *		to 1.6 mm	DAkks-DKD-R 4-3 part 11.3:2010	1 μm	
Lever gauges for external measurements *	0 mm	to 70 mm	DAkks-DKD-R 4-3 part 12.1:2010	$7\text{ }\mu\text{m} + 10 \cdot 10^{-6} \cdot l$	up to a probe length of 200 mm

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-01

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range		Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Lever gauges for internal measurements *	2.5 mm to	80 mm	DAkkS-DKD-R 4-3 part 13.1:2010	$7 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	l = measured length
Thickness gauges *	0 mm to	30 mm	DAkkS-DKD-R 4-3 part 12.1:2010	$7 \mu\text{m} + 10 \cdot 10^{-6} \cdot l$	l = measured length up to a measuring depth of 200 mm
Bore gauges with two-point contact *	1 mm to	3 mm	VDI/VDE/DGQ 2618 part 13.2:2005 (image 1)	0.8 μm	range of application: with gauge slider $d = 1.75 \text{ mm}$ to $d = 25 \text{ mm}$
			VDI/VDE/DGQ 2618 part 13.2:2005 (image 2)	0.8 μm	range of application: to $d = 300 \text{ mm}$
				1.2 μm	range of application: $d > 300 \text{ mm}$ to $d = 600 \text{ mm}$
			VDI/VDE/DGQ 2618 part 13.2:2005 (image 3)	0.8 μm	range of application: plug gauge to $d = 100 \text{ mm}$
Height gauges *	0 mm to	1000 mm	VDI/VDE/DGQ 2618 part 16.1:2009	$1.5 \mu\text{m} + 3 \cdot 10^{-6} \cdot L$	L = measured length
Deviation from straightness and Perpendicularity	to	30 μm	to 800 mm lead length	$2.5 \mu\text{m} + 1 \cdot 10^{-6} \cdot l_z$	l_z = lead length
90° Squares Perpendicularity	to	30 μm	Annex F/12 V3:2010 to 750 mm leg length	$2 \mu\text{m} + 2 \cdot 10^{-6} \cdot l_z$	l_z = length of form respectively position embodiment
Deviation from straightness and parallelism	to	30 μm	to 750 mm length	$1 \mu\text{m} + 2 \cdot 10^{-6} \cdot l_z$	
Flatness deviation	to	30 μm	to 750 mm edge length	$1 \mu\text{m} + 2 \cdot 10^{-6} \cdot l_z$	
Deviation from flatness Horizontal flatness standard, e.g. surface plates as per DIN 876:1984	to	50 μm	Annex F/13 V2:2010 to 2 m edge length electronic inclination measuring	$1 \mu\text{m} + 1 \cdot 10^{-6} \cdot l$	l = longest edge length of the measuring standard For calibration in the DAkkS calibration labora- tory, the uncertainty increases starting by an edge length
Deviation from straightness Horizontal flatness standard, e.g. surface plates as per DIN 876:1984	to	50 μm	Annex F/13 V2:2010 to 5 m edge length electronic inclination measuring	$1 \mu\text{m} + 1 \cdot 10^{-6} \cdot l$	$l > 1 \text{ m}$ by a factor of 1.2

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-01
Permanent Laboratory
Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Thread gauges * (single-start and multi-start cylindrical and conical external and internal threads with straight flanks, symmetrical and asymmetrical profile)				
External thread	3 mm to 90 mm	Scanning method DAkkS-DKD-R 4-3 part 4.8:2010, Option 1 to option 4 (Specifying the thread angle α)		
Simple pitch diameter	Nominal diameter		2.5 μm	
Outside diameter			2 μm	
Core diameter or recess diameter			5 μm	
Lead or pitch	0.5 mm to 8 mm		1 μm	
Thread angle α	$\geq 27^\circ$		$(1.2 + 3 \text{ mm} / l_F)'$, but not lower at $6'$	
Internal thread	3 mm to 100 mm	Scanning method DAkkS-DKD-R 4-3 part 4.9:2010, Option 1 to option 4 (Specifying the thread angle α)		
Simple pitch diameter	Nominal diameter		2.5 μm	
Outside diameter or recess diameter			5 μm	
Core diameter			2 μm	
Lead or pitch	0.5 mm to 8 mm		1 μm	
Thread angle α	$\geq 27^\circ$		$(1.2 + 3 \text{ mm} / l_F)'$, but not lower at $6'$	
Thread plug gauge * Simple pitch diameter	1.4 mm to 150 mm	DAkkS-DKD-R 4-3 part 4.8:2010, Option 1 (Three wire procedure)	$2.5 \mu\text{m} + 7.5 \cdot 10^{-6} \cdot d$	d = measured diameter P_h = lead $P_h \geq 0.3 \text{ mm}$ to $\leq 6 \text{ mm}$
Torque *				
Hand torque assembly tools	1 N·m to 1000 N·m	DIN EN ISO 6789-2:2017	0.5 %	only operated torque tools
Torque wrench calibration devices	4 N·m to 1000 N·m	DAkkS-DKD-R 3-8:2010	0.2 %	
Pressure *				
gauge pressure p_e	1 bar to 700 bar	DIN EN 837:1997 DKD-R 6-1:2014	0.2 bar	Pressure medium: oil
	$\geq 700 \text{ bar}$ to 800 bar		0.5 bar	
gauge pressure p_e	1 bar to 30 bar		0.01 bar	Pressure medium: gas

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-01

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Weighing instruments * Nonautomatic weighing instruments	to ≤ 50 kg	EURAMET Calibration Guide No. 18 version 4.0 (11/2015)	$1,2 \cdot 10^{-5}$	with weights OIML R 111-1:2004 according to the class F1
Material testing machines (MTM) Hardness (MTM) * Hardness Testers according to hardness scales Shore A, AO and D	0 Shore to 100 Shore	DIN ISO 7619-1:2012 DIN ISO 18898:2014	1 Shore	Direct measurement with reference standards for travel and power. Optical calibration of the geometrical measurements with optical and tactile coordinate measuring machines
Measuring range	to 2.5 mm		6 µm	
Diameter, radii, lengths	to 27 mm		3.5 µm	
Area	to 600 mm ²		5 µm ²	
Angle	28° to 37°		0.1°	
Elastic force	0 N to 44.5 N		0.5 % of final value	
Shore A, AO and D Measuring path standard	0.5 mm to 2.5 mm	Annex F/34 V1:2014	0.8 µm	
Temperature quantities * Temperature indicators for thermocouples	-200 °C to 1300 °C	DAkkS-DKD-R 5-5:2010	0.5 K	Simulation of the thermo-electric on multifunction generator (input in temperature units (° C)) Electric consideration of the reference junction
Temperature indicators for resistance thermo-meters with sensor PT100	-100 °C to 800 °C		0.2 K	Simulation of the resistance value on multifunction calibrator (input in temperature units (° C))
DC and low frequency quantities DC voltage Measuring instruments	0 mV to 200 mV > 0.2 V to 2 V > 2 V to 20 V > 20 V to 200 V > 200 V to 1000 V		$2.8 \mu\text{V} + 16 \cdot 10^{-6} \cdot U$ $2.8 \mu\text{V} + 10 \cdot 10^{-6} \cdot U$ $28 \mu\text{V} + 10 \cdot 10^{-6} \cdot U$ $0.28 \text{ mV} + 15 \cdot 10^{-6} \cdot U$ $2.8 \text{ mV} + 15 \cdot 10^{-6} \cdot U$	U = measuring value
Sources	0 V to 0.1 V > 0.1 V to 1 V > 1 V to 10 V > 10 V to 100 V > 100 V to 1000 V		$0.77 \mu\text{V} + 2.6 \cdot 10^{-6} \cdot U$ $0.88 \mu\text{V} + 4.8 \cdot 10^{-6} \cdot U$ $6.6 \mu\text{V} + 5.4 \cdot 10^{-6} \cdot U$ $98 \mu\text{V} + 7.2 \cdot 10^{-6} \cdot U$ $1.4 \text{ mV} + 7.3 \cdot 10^{-6} \cdot U$	

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-01

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
DC current Measuring instruments	0 μ A to 200 μ A > 0.2 mA to 2 mA > 2 mA to 20 mA > 20 mA to 200 mA > 0,2 A to 2 A > 2 A to 20 A		16 nA + 0.1 $\cdot 10^{-3} \cdot I$ 57 nA + 50 $\cdot 10^{-6} \cdot I$ 0.6 μ A + 45 $\cdot 10^{-6} \cdot I$ 2.4 μ A + 0.12 $\cdot 10^{-3} \cdot I$ 0.26 mA + 87 $\cdot 10^{-6} \cdot I$ 0.55 mA + 0.34 $\cdot 10^{-3} \cdot I$	I = measuring value
DC current sources	10 μ A to 100 μ A > 0.1 mA to 1.0 mA > 1.0 mA to 10 mA > 10 mA to 100 mA > 0.1 A to 1.0 A > 1 A to 10 A > 10 A to 30 A		0.46 nA + 9.5 $\cdot 10^{-6} \cdot I$ 4.9 nA + 9.3 $\cdot 10^{-6} \cdot I$ 0.05 μ A + 11 $\cdot 10^{-6} \cdot I$ 0.71 μ A + 35 $\cdot 10^{-6} \cdot I$ 15 μ A + 0.17 $\cdot 10^{-3} \cdot I$ 0.4 mA + 0.42 $\cdot 10^{-3} \cdot I$ 5 mA + 0.58 $\cdot 10^{-3} \cdot I$	I = measuring value
DC current clamps	0.2 A to < 10 A 10 A to 100 A > 100 A to 1000 A		10 mA + 2 $\cdot 10^{-3} \cdot I$ 0.1 A + 2 $\cdot 10^{-3} \cdot I$ 0.8 A + 2.5 $\cdot 10^{-3} \cdot I$	I = measuring value with current coil with 2, 10 and 50 windings
AC voltage Measuring instruments	0.02 V to 0.2 V > 0.2 V to 2 V > 2 V to 20 V > 20 V to 200 V > 200 V to 1000 V	45 Hz to 999 Hz 1 kHz to 19.999 kHz 45 Hz to 999 Hz 1 kHz to 19.999 kHz 45 Hz to 999 Hz 1 kHz to 19.999 kHz 45 Hz to 999 Hz 1 kHz to 10 kHz 45 Hz to 999 Hz 1 kHz to 10 kHz	24 μ V + 0.2 $\cdot 10^{-3} \cdot U$ 36 μ V + 0.25 $\cdot 10^{-3} \cdot U$ 0.15 mV + 0.21 $\cdot 10^{-3} \cdot U$ 0.2 mV + 0.3 $\cdot 10^{-3} \cdot U$ 1.1 mV + 0.22 $\cdot 10^{-3} \cdot U$ 1.8 mV + 0.3 $\cdot 10^{-3} \cdot U$ 14 mV + 0.2 $\cdot 10^{-3} \cdot U$ 18 mV + 0.3 $\cdot 10^{-3} \cdot U$ 0.07 V + 0.26 $\cdot 10^{-3} \cdot U$ 0.13 V + 0.35 $\cdot 10^{-3} \cdot U$	U = measuring value
AC voltage Sources	10 mV to 100 mV > 0.1 V to 1.0 V > 1.0 V to 10 V > 10 V to 100 V > 100 V to 1000 V	40 Hz to 200 Hz > 200 Hz to 2 kHz > 2 kHz to 20 kHz 40 Hz to 200 Hz > 200 Hz to 2 kHz > 2 kHz to 20 kHz 40 Hz to 200 Hz > 200 Hz to 2 kHz > 2 kHz to 20 kHz 40 Hz to 200 Hz > 200 Hz to 2 kHz > 2 kHz to 20 kHz	11 μ V + 0.33 $\cdot 10^{-3} \cdot U$ 11 μ V + 0.27 $\cdot 10^{-3} \cdot U$ 13 μ V + 0.34 $\cdot 10^{-3} \cdot U$ 0.07 mV + 0.23 $\cdot 10^{-3} \cdot U$ 0.06 mV + 0.22 $\cdot 10^{-3} \cdot U$ 0.07 mV + 0.94 $\cdot 10^{-3} \cdot U$ 0.7 mV + 0.24 $\cdot 10^{-3} \cdot U$ 0.7 mV + 0.23 $\cdot 10^{-3} \cdot U$ 0.73 mV + 0.93 $\cdot 10^{-3} \cdot U$ 10 mV + 0.3 $\cdot 10^{-3} \cdot U$ 7.6 mV + 0.28 $\cdot 10^{-3} \cdot U$ 8 mV + 0.88 $\cdot 10^{-3} \cdot U$ 83 mV + 0.48 $\cdot 10^{-3} \cdot U$ 52 mV + 0.69 $\cdot 10^{-3} \cdot U$ 58 mV + 1.4 $\cdot 10^{-3} \cdot U$	

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-01

Permanent Laboratory

Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
AC current Measuring instruments	20 μ A to 200 μ A > 0.2 mA to 2 mA > 2 mA to 20 mA > 20 mA to 200 mA > 0.2 A to 2 A > 2 A to 20 A	45 Hz to 1 kHz 45 Hz to 1 kHz 45 Hz to 1 kHz 45 Hz to 1 kHz 45 Hz to 1 kHz 45 Hz to 100 Hz	$0.23 \mu\text{A} + 1.1 \cdot 10^{-3} \cdot I$ $0.23 \mu\text{A} + 1.1 \cdot 10^{-3} \cdot I$ $2.3 \mu\text{A} + 1.1 \cdot 10^{-3} \cdot I$ $23 \mu\text{A} + 1.1 \cdot 10^{-3} \cdot I$ $0.23 \text{ mA} + 1.1 \cdot 10^{-3} \cdot I$ $2.3 \text{ mA} + 1.1 \cdot 10^{-3} \cdot I$	I = measuring value
AC current clumps	0.2 A to < 10 A 10 A to 100 A > 100 A to 500 A > 500 A to 1000 A	50 Hz 50 Hz 50 Hz 50 Hz	$10 \text{ mA} + 2 \cdot 10^{-3} \cdot I$ $0.1 \text{ A} + 2 \cdot 10^{-3} \cdot I$ $0.4 \text{ A} + 2.5 \cdot 10^{-3} \cdot I$ $0.8 \text{ A} + 2.5 \cdot 10^{-3} \cdot I$	I = measuring value with current coil with 2, 10 and 50 windings
AC current sources	0.1 mA to 1.0 mA > 1.0 mA to 10 mA > 10 mA to 100 mA > 0.1 A to 1.0 A > 1.0 A to 10 A > 10 A to 30 A	40 Hz to 1 kHz 40 Hz to 1 kHz 40 Hz to 1 kHz 40 Hz to 1 kHz 40 Hz to 1 kHz 40 Hz to 1 kHz	$0.17 \mu\text{A} + 0.35 \cdot 10^{-3} \cdot I$ $1.4 \mu\text{A} + 0.37 \cdot 10^{-3} \cdot I$ $14 \mu\text{A} + 0.37 \cdot 10^{-3} \cdot I$ $0.17 \text{ mA} + 0.49 \cdot 10^{-3} \cdot I$ $4.6 \text{ mA} + 0.85 \cdot 10^{-3} \cdot I$ $14 \text{ mA} + 0.87 \cdot 10^{-3} \cdot I$	I = measuring value
DC resistance Measuring instruments 4-wire connection	0 Ω 0.1 Ω 1 Ω 10 Ω 100 Ω 1 k Ω 10 k Ω 100 k Ω		5.8 m Ω 5.8 m Ω 6 m Ω 7 m Ω 12 m Ω 93 m Ω 0.93 Ω 9.3 Ω	Fixed resistors
DC resistances 4-wire connection	0 Ω to 1 Ω > 1 Ω to 10 Ω > 10 Ω to 100 Ω > 100 Ω to 1 k Ω > 1 k Ω to 10 k Ω > 10 k Ω to 100 k Ω		$0.23 \text{ m}\Omega + 2 \cdot 10^{-6} \cdot R$ $0.23 \text{ m}\Omega + 4 \cdot 10^{-6} \cdot R$ $0.27 \text{ m}\Omega + 9 \cdot 10^{-6} \cdot R$ $1.6 \text{ m}\Omega + 8.3 \cdot 10^{-6} \cdot R$ $15 \text{ m}\Omega + 10 \cdot 10^{-6} \cdot R$ $0.28 \Omega + 9.7 \cdot 10^{-6} \cdot R$	R = measuring value
DC resistance Measuring instruments 2-wire connection	1 M Ω 10 M Ω 100 M Ω 1000 M Ω		0.16 k Ω 4.6 k Ω 0.6 M Ω 14 M Ω	Fixed resistors
DC resistances 2-wire connection	> 100 k Ω to 1 M Ω > 1 M Ω to 10 M Ω		$5.8 \Omega + 10 \cdot 10^{-6} \cdot R$ $0.1 \text{ k}\Omega + 17 \cdot 10^{-6} \cdot R$	R = measuring value
Time and frequency Frequency Measuring instruments	10 Hz 100 Hz 1 kHz 10 kHz 20 kHz 50 kHz 100 kHz 1 MHz 10 MHz		$23 \cdot 10^{-6} \cdot F$	F = measuring value

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-01
Permanent Laboratory
Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Revolution speed Revolution counter optical	240 min ⁻¹ bis 60000 min ⁻¹		0.1 min ⁻¹ + 18 · 10 ⁻⁶ · <i>n</i>	direct optical excitation <i>n</i> = measuring value

On-site Calibration
Calibration and Measurement Capabilities (CMC)

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Length				
Deviation from flatness Horizontal flatness standard, e.g. surface plates as per DIN 876:1984	to 50 µm	Annex F/13 V2:2010 to 2 m edge length electronic inclination measuring	1 µm + 1 · 10 ⁻⁶ · <i>l</i>	<i>l</i> = longest edge length of the measuring standard
Deviation from straightness Horizontal flatness standard, e.g. surface plates as per DIN 876:1984	to 50 µm	Annex F/13 V2:2010 to 5 m edge length electronic inclination measuring	1 µm + 1 · 10 ⁻⁶ · <i>l</i>	
Height gauges *	0 mm to 600 mm	VDI/VDE/DGQ 2618 part 16.1:2009	2.5 µm + 5 · 10 ⁻⁶ · <i>L</i>	<i>L</i> = measured length

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of *k* = 2 unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-01

On-site Calibration

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Coordinate measuring technology * Measuring projectors Measuring microscopes	Devices featuring a measuring surface with a face diagonal ≤ 530 mm	Calibration of metro-logical characteristics according to DAkkS-DKD-R 4-3 part 18.1:2010, and the folloing standards and gudelines DIN EN ISO 10360 VDI/VDE 2617		
		Determination of probing error $PS-ID(OT)$ with a graduated scale made of glass according to VDI/VDE 2617 part 6.1:2007	0.4 μ m	Measuring projectors and measuring Microscopes with visual probing with crosshairs or electronic edge detection
		The error of indication for size measurement $E-ID(OT)$ and $E-2D(OT)$ is determined with a graduated scale made of glass according to VDI/VDE 2617 part 6.1:2007	$0.5 \mu\text{m} + 0.5 \cdot 10^{-6} \cdot L$	L = measured length
Weighing instruments * Nonautomatic weighing instruments	to ≤ 50 kg	EURAMET Calibration Guide No. 18 version 4.0 (11/2015)	$1 \cdot 10^{-5}$	with weights OIML R 111-1:2004 according to the class F1
Torque * Hand torque assembly tools	1 N·m to 1000 N·m	DIN EN ISO 6789-2:2017	0.5 %	only operated torque tools

Mobile Laboratory

Calibration and Measurement Capabilities (CMC)				
Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement ¹⁾	Remarks
Torque * Hand torque assembly tools	1 N·m to 1000 N·m	DIN EN ISO 6789-2:2017	0.5 %	only operated torque tools

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.

Annex to the accreditation certificate D-K-15089-01-01**Abbreviations used:**

CMC	Calibration and measurement capabilities (Kalibrier- und Messmöglichkeiten)
DIN	Deutsches Institut für Normung e.V.
DAkkS-DKD-R	Guideline: Deutsche Akkreditierungsstelle GmbH
DKD-R	Guideline: Deutscher Kalibrierdienst (DKD), issued by Physikalisch-Technische Bundesanstalt
EURAMET	European Association of National Metrology Institutes
VDI/VDE 2617	Guideline: Accuracy of coordinate measuring machines
VDI/VDE/DGQ 2618	Guideline: Inspection of measuring and test equipment
VDE	Verband der Elektrotechnik, Elektronik und Informationstechnik
VDI	Verein Deutscher Ingenieure

¹⁾ The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of $k = 2$ unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.